

1.14774.0001

MColortest™

Nitrite Test

NO₂⁻

1. Method

Determination with color-disk comparator

In acidic solution nitrite ions react with sulfanilic acid to form a diazonium salt, which in turn reacts with N-(1-naphthyl)ethylenediamine dihydrochloride to form a red-violet azo dye. The nitrite concentration is measured **semiquantitatively** by visual comparison of the color of the measurement solution with the color fields of a color disk.

2. Measuring range and number of determinations

Measuring range / color-scale graduation ¹⁾	Number of determinations
0.1 - 0.2 - 0.4 - 0.6 - 1.0 - 1.8 - 3.0 - 6.0 - 10 mg/l NO ₂ ⁻	400
0.03 - 0.06 - 0.12 - 0.18 - 0.30 - 0.55 - 0.9 - 1.8 - 3.0 mg/l NO ₂ -N	

¹⁾ for conversion factors see section 8

3. Applications

Sample material:

Groundwater and surface water, seawater
Drinking water and mineral water
Waters from aquaculture
Boiler and boiler feed water, cooling water
Wastewater and electroplating wastewater
Food after appropriate sample pretreatment
Soils after appropriate sample pretreatment

4. Influence of foreign substances

This was checked in solutions containing 1 and 0 mg/l NO₂⁻. The determination is not yet interfered with up to the concentrations of foreign substances given in the table.

Concentrations of foreign substances in mg/l or %					
Ca ²⁺	1000	Hg ²⁺	100	SiO ₃ ²⁻	1000
Cd ²⁺	1000	Mg ²⁺	1000	Zn ²⁺	1000
CN ⁻	1000	Mn ²⁺	1000	EDTA	1000
Cr ³⁺	100	NH ₄ ⁺	1000	Reducing agents (ascorbic acid, sulfite)	10
Cr ₂ O ₇ ²⁻	1	Pb ²⁺	1000	NaCl	20 %
Cu ²⁺	100	PO ₄ ³⁻	1000	NaNO ₃	20 %
Fe ³⁺	1	S ²⁻	10	Na ₂ SO ₄	15 %

5. Reagents and auxiliaries

Please note the warnings on the packaging materials!

The test reagent is stable up to the date stated on the pack when stored closed at +15 to +25 °C.

Package contents:

3 bottles of reagent NO₂-1
1 graduated 6-ml plastic syringe
2 test tubes with screw caps
1 colour-disk comparator

Other reagents and accessories:

MQuant™ Nitrite Test, Cat. No. 110007,
measuring range 2 - 80 mg/l NO₂⁻ (0.6 - 24 mg/l NO₂-N)
MColorpHast™ Universal indicator strips pH 0 - 14, Cat. No. 109535
MColorpHast™ pH-indicator strips pH 0 - 6.0, Cat. No. 109531
Sulfuric acid 0.5 mol/l TitriPUR®, Cat. No. 109072
Sodium hydroxide solution 1 mol/l TitriPUR®, Cat. No. 109137
Nitrite standard solution CertiPUR®, 1000 mg/l NO₂⁻, Cat. No. 119899

MColortest™ Flat-bottomed tubes with screw caps for MColortest™ with color-disk comparator (12 pcs), Cat. No. 117988

Refill pack:

Cat. No. 118463

Nitrite Test

Refill pack for 114774, 114424, and 114408

(Reagent **without technical accessories** for the number of determinations stated in section 2)

6. Preparation

- Analyze immediately after sampling.
- Check the nitrite content with the MQuant™ Nitrite Test. Samples containing more than 10 mg/l NO₂⁻ must be diluted with distilled water.
- The pH must be within the range 2 - 10.** Adjust, if necessary, with sulfuric acid.
- Filter strongly turbid samples.

7. Procedure

	Measurement sample right-hand tube (A) behind the color disk	Blank left-hand tube (B) behind the color disk	
Pretreated sample (15 - 25 °C) Reagent NO ₂ -1	6 ml 1 level blue microspoon (in the cap of the NO ₂ -1 bottle)	6 ml -	Inject into the test tube with the syringe. Add, close the tube, and shake vigorously until the reagent is completely dissolved. The pH must be within the range 2.0 - 2.5. Check with pH-indicator strips. Adjust the pH, if necessary, with sodium hydroxide solution or sulfuric acid.
Leave to stand for 3 min (reaction time).			
Hold the comparator to the light, keeping it upright, and rotate the disk until the closest possible color match is achieved between the two large windows. Read off the result in mg/l NO ₂ ⁻ shown in the small window.			

Notes on the measurement:

- The color of the measurement solution remains stable for at least 60 min after the end of the reaction time stated above.
- Turbidity in the measurement solution makes the color comparison more difficult.
- If the color of the measurement solution is equal to or more intense than the darkest color on the scale, repeat the measurement using **fresh**, diluted samples until a value of less than 10 mg/l NO₂⁻ is obtained.

Concerning the result of the analysis, the dilution (see also section 6) must be taken into account:

$$\text{Result of analysis} = \text{measurement value} \times \text{dilution factor}$$

8. Conversions

Units required	=	units given	x	conversion factor
mg/l NO ₂ -N		mg/l NO ₂ ⁻		0.304
mg/l NO ₂ ⁻		mg/l NO ₂ -N		3.28

9. Method control

To check test reagent, measurement device, and handling: Dilute the nitrite standard solution with distilled water to 1.0 mg/l NO₂⁻ and analyze as described in section 7.

Additional notes see under www.qa-test-kits.com.

10. Notes

- Reclose the reagent bottle immediately after use.
- Rinse the test tubes and the syringe **with distilled water only**.
- Information on disposal can be obtained at www.disposal-test-kits.com.**

Merck KGaA, 64271 Darmstadt, Germany,
Tel. +49(0)6151 72-2440
www.analytical-test-kits.com

EMD Millipore Corporation, 290 Concord Road,
Billerica, MA 01821, USA, Tel. +1-978-715-4321

